


41. At M-step of EM-algorithm for two classes, if the first class expected responsibility matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, then the prior probability of the first class is
 (a) 0.25 (b) 0.375 (c) 0.9 (d) none of the given
42. At M-step of EM-algorithm for two classes, if the first class responsibility expected matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, then the prior probability of the second class is
 (a) 0.25 (b) 0.375 (c) 0.9 (d) none of the given
43. At M-step, if the first class responsibility matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, and the image is $Y = [1 \ 0; 0 \ 1]$, then the mean of the first class is
 (a) 0.8 (b) 1.2 (c) 1.5 (d) 2.134
44. At M-step, if the first class responsibility matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, and the image is $Y = [1 \ 0; 0 \ 1]$ then the mean of the second class is
 (a) 2.5 (b) 1.92 (c) 1.2 (d) 0.2
45. In MRI, the RF coil.....
 (a) transmits excitation signal (b) receives body magnetic response (c) aligns protons in the filed direction (d) both (a)&(b)
46. The mass number of the following nucleus represent an MRI active nucleus:
 (a) 18 (b) 12 (c) 11 (d) either (a) or (b)
47. Essentially all MRI is aimaging
 (a) water's proton (b) nitrogen (c) water's oxygen (d) carbon's proton
48. During excitation, the net magnetization vector flips with an angle that is a function of
 (a) strength of the RF pulse field (b) duration of the RF pulse (c) strength of the main magnet field (d) both (a)&(b)
49. Regularly, brain white matter and grey matter have different MRI relaxation time(s)
 (a) T1 (b) T2 (c) either (a) or (b) (d) both (a)&(b)
50. studies the activated area of the brain after certain stimuli
 (a) Tagged MRI (b) Diffusion imaging (c) Functional MRI (d) T1-weighted MRI

ALL BEST WISHES Assistant Prof. Ahmed Elnakib



TOTAL of 50 points, each is one point, select one choice, only one choice is correct



1. CT imaging is associated with the interaction ofwaves with body
 (a) x-ray (b) ultrasonic (c) magnetic field (d) gamma-ray
2. The scan of a..... is much more safe than a CT scan of the abdomen
 (a) dental teeth (b) colon CT (c) stomach CT (d) chest CT
3. If the attenuation coefficient of water is 0.2 cm^{-1} , then its Hounsfield unit is
 (a) 0.2 (b) 0 (c) positive (d) negative
4. The relation between the transmitted x-ray intensity with distance inside tissue is.....
 (a) linear (b) quadrature (c) cubic (d) exponential
5. In Fig. A, applying a first step vertical ART correction, set $A = \dots \text{cm}^{-1}$
 (a) 0.5 (b) 0.25 (c) 0.125 (d) none of the given
6. In Fig. A, applying a 2nd step horizontal ART correction, after vertical, set $B = \dots \text{cm}^{-1}$
 (a) 0.11 (b) -0.08 (c) 0.08 (d) 0.27
7. In Fig. A, applying a third step diagonal (main and reverse) ART correction, after vertical and horizontal, respectively, set $C = \dots \text{cm}^{-1}$
 (a) 0.1 (b) -0.07 (c) 0.07 (d) 0.24
8. If the sum of absolute minimum acceptable errors of one cycle (including vertical, horizontal, and diagonal steps) is 0.01, then it needcycles to terminate the ART algorithm
 (a) zero (b) one (c) two (d) undetermined
9. According to Fig. A, Pixel D is a.....tissue
 (a) water (b) fat (c) soft (d) lung
10. To select a parametric probabilistic intensity model for image classes, one should look at..
 (a) image histogram (b) image resolution (c) EM components (d) prior values
11. A Gaussian probability density function (pdf) is described by the parameter(s).....
 (a) prior values (b) means and variances (c) means (d) both (a) & (b)
12. According to the Bayes decision theory, the apriori increases linearly with.....
 (a) prior value (b) conditional probability (c) evidence (d) both (a) & (b)

13. If the number of object pixels=200 in a 100x100 training two-class image, then the prior probability of the other class (background) is
- (a) 200 (b) 0.02 (c) 0.98 (d) none of the given
14. Let Y be the grey-level image, then the Bayes decision theory selects class 1 ($x=0$) if.....
- (a) $p(Y|x=0) > p(Y|x=1)$ (b) $p(X=0) > p(X=1)$ (c) $p(x=0|Y) > p(x=1|Y)$ (d) none of the given
15. If a 5-fold cross validation is applied on a data of total size of 1000 record, then the training data should be of size.....record
- (a) 800 (b) 500 (c) 200 (d) 5
16. If a leave-one-out cross validation is applied on a data of total size of 1000 record, then one should apply a number of experiments
- (a) 1000 (b) 999 (c) 200 (d) 1
17. To more accurately describe an accuracy to be 0.8, three experiments accuracies are....
- (a) 0.7,0.8,0.9 (b) 0.75,0.8,0.85 (c) 0.6,0.8,1 (d) any answer
18. To extract deep learning features from a 50x50 image using a convolution with a kernel of size 5x5, the output feature map is of size....
- (a) 50x50 (b) 45x45 (c) 46x46 (d) 5x5
19. The job of convolution with a kernel of size 3x3 with all values of 1/9 is to apply
- (a) smoothing (b) sharpening (c) high pass filter (d) both (a)&(b)
20. To extract deep learning features from a 3x3 image $Y=[0 \ 1 \ 1; 1 \ 0 \ 1; 1 \ 1 \ 0]$ using a convolution with a kernel of size 3x3 with all values of 1/9, then the first output pixel is.....
- (a) 1/9 (b) 2/3 (c) 1 (d) 3
21. If a bat hears an echo 3.4 second after it makes a chirp of speed 340 m/sec, how far away is the object?
- (a) 100m (b) 50m (c) 200m (d) 578m
22. High ultrasonic imaging frequencies is suitable for imaging
- (a) small size objects (b) surface structures (c) deep structures (d) both (a)&(b)
23. In amplitude-mode ultrasonic (US) scan, the.....decreases as the amplitude increases
- (a) tissue reflectivity (b) object depth (c) transducer size (d) both (a)&(b)
24. Compound imaging (sonoCT) is an enhanced scan
- (a) A-mode (b) B-mode (c) M-mode (d) none of the given answers
25. If a blood cell is moving with a velocity u m/s towards a transducer that submits an ultrasonic wave with a frequency ν and speed C , then the reflected wave has
- (a) same speed C (b) different frequency ν (c) different speed C' (d) both (a)&(b)
26. Continuous wave Doppler affords.....measurements of blood speed
- (a) high speed value (b) large area (c) small area (d) both (a)&(b)


27. If a blood cell is moving with a velocity 500 m/s away from a transducer that submits an ultrasonic wave with 100 MHz and 1500 m/s, then the reflected frequency is.... MHz
- (a) 75 (b) 100 (c) 150 (d) none of the given
28. If a blood cell is moving with a velocity 0 m/s and a transducer submits an ultrasonic wave with 100 MHz and 1500 m/s, then the reflected frequency is.... MHz
- (a) 75 (b) 100 (c) 150 (d) none of the given
29. Color Doppler superimpose blood velocity map on top of a(n)..... scan
- (a) A-mode (b) B-mode (c) M-mode (d) none of the given
30. Forms of ionized radiation include....
- (a) MRI waves (b) ultrasonic wave (c) microwave (d) gamma-rays
31. Positron emission tomography basically used a camera to detect
- (a) positrons (b) gamma-rays (c) alpha particles (d) either (a) or (b)
32. A radioactivity is dropped by one quarter after travelling 2cm in tissue and by one more quarter after travelling a more 1cm in the tissue, then.....
- (a) half life=2cm (b) half life=3cm (c) half value layer=2cm (d) half value layer=3cm
33. Tumor therapy applications would use ...particles due its half value layer
- (a) alpha/small (b) alpha/large (c) positron/small (d) positron/large
34. A cyclotron is used onsite for PET scan as the positron-emitting radionuclide.....
- (a) has short half time (b) is synthesized (c) exists naturally (d) both (a)&(b)
35. To build a prior shape model, a registration step is done to..... training images
- (a) maximize similarities between (b) enhance visual appearance of (c) minimize similarities between (d) both (a)&(b)
36. If among 100 training images, the pixel ($x=5, y=5$) is an object for 20 images, then the probabilistic shape value for this pixel to belong to the background is
- (a) 20/100 (b) 80/100 (c) (5x5)/100 (d) none of the given
37. If the pixel ($x=5, y=5$) in the training image belong to the object edge, then the distance map shape value for this pixel is
- (a) 0 (b) 1 (c) -1 (d) none of the given
38. If the pixel ($x=50, y=50$) in the training image is deep inside the object, then the distance map shape value for this pixel is
- (a) 0 (b) positive (c) negative (d) 50x50
39. If the pixel ($x=50, y=50$) is an edge point in the training image, then the distance map shape value for the pixel ($x=51, y=51$) may be
- (a) 0 (b) root of 2 (c) -2 (d) none of the given
40. The M-step of the EM algorithm computes the updated expected value of the....probability
- (a) log likelihood (b) prior (c) conditional (d) both (a)&(b)

42. Essentially all MRI is a.....imaging
 (a) water's proton (b) nitrogen (c) water's oxygen (d) carbon's proton
43. During excitation, the net magnetization vector flips with an angle that is a function of
 (a) strength of the RF pulse field (b) duration of the RF pulse (c) strength of the main magnet field (d) both (a)&(b)
44. Regularly, brain white matter and grey matter have different MRI relaxation time(s)
 (a) T1 (b) T2 (c) either (a) or (b) (d) both (a)&(b)
55. studies the activated area of the brain after certain stimuli
 (a) Tagged MRI (b) Diffusion imaging (c) Functional MRI (d) T1-weighted MRI
46. At M-step of EM-algorithm for two classes, if the first class expected responsibility matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, then the prior probability of the first class is
 (a) 0.25 (b) 0.375 (c) 0.9 (d) none of the given
47. At M-step of EM-algorithm for two classes, if the first class responsibility expected matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, then the prior probability of the second class is
 (a) 0.25 (b) 0.375 (c) 0.9 (d) none of the given
48. At M-step, if the first class responsibility matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, and the image is $Y = [1 \ 0; 0 \ 1]$, then the mean of the first class is
 (a) 0.8 (b) 1.2 (c) 1.5 (d) 2.134
49. At M-step, if the first class responsibility matrix $\pi_1 = [0.8 \ 0.4; 0.3 \ 0]$, and the image is $Y = [1 \ 0; 0 \ 1]$ then the mean of the second class is
 (a) 2.5 (b) 1.92 (c) 1.2 (d) 0.2
50. The M-step of the EM algorithm computes the updated expected value of the....probability
 (a) log likelihood (a) log likelihood (a) log likelihood (a) log likelihood

ALL BEST WISHESAssistant Prof. Ahmed Elnakhib



TOTAL of 50 points, each is one point, select one choice, only one choice is correct



- Nuclear imaging is associated with the interaction ofwaves with body
 (a) x-ray (b) ultrasonic (c) magnetic field (d) gamma-ray
- In Fig. A, applying a first step vertical ART correction, set A=.....cm⁻¹
 (a) 0.5 (b) 0.25 (c) 0.125 (d) none of the given
- In Fig. A, applying a 2nd step horizontal ART correction, after vertical, set B=.....cm⁻¹
 (a) 0.11 (b) -0.08 (c) 0.08 (d) 0.27
- In Fig. A, applying a third step diagonal (main and reverse) ART correction, after vertical and horizontal, respectively, set C=.....cm⁻¹
 (a) 0.1 (b) -0.07 (c) 0.07 (d) 0.24
- If the sum of absolute minimum acceptable errors of one cycle (including vertical, horizontal, and diagonal steps) is 0.01, then it needcycles to terminate the ART algorithm
 (a) zero (b) one (c) two (d) undetermined
- According to Fig. A, Pixel D is a.....tissue
 (a) water (b) fat (c) soft (d) lung
- The scan of a..... is much more safe than a CT scan of the abdomen
 (a) dental teeth (b) colon CT (c) stomach CT (d) chest CT
- If the attenuation coefficient of water is 0.2 cm⁻¹, then its Hounsfield unit is
 (a) 0.2 (b) 0 (c) positive (d) negative
- The relation between the transmitted x-ray intensity with distance inside tissue is.....
 (a) linear (b) quadrature (c) cubic (d) exponential
- To select a parametric probabilistic intensity model for image classes, one should look at.
 (a) image resolution (b) image histogram (c) EM components (d) prior values
- A Gaussian probability density function (pdf) is described by the parameter(s).....
 (a) prior values (b) means (c) means and variances (d) both (a) & (b)
- According to the Bayes decision theory, the aposteriori decreases linearly with.....
 (a) prior value (b) conditional probability (c) evidence (d) both (a) & (b)

MODEL BOnly answers in the electronic answer form will be graded

13. If the number of object pixels=200 in a 100x100 training two-class image, then the prior probability of the **other class (background)** is
- (a) 200 (b) 0.98 (c) 0.02 (d) none of the given
14. Let Y be the grey-level image, then the Bayes decision theory selects class 1 ($x=0$) if.....
- (a) $p(Y|x=0) > p(Y|x=1)$ (b) $p(x=0|Y) > p(x=1|Y)$ (c) $p(X=0) > p(X=1)$ (d) none of the given
15. To extract deep learning features from a 50x50 image using a convolution with a kernel of size 5x5, the output feature map is of size....
- (a) 50x50 (b) 45x45 (c) 46x46 (d) 5x5
16. The job of convolution with a kernel of size 3x3 with all values of 1/9 is to apply
- (a) smoothing (b) sharpening (c) high pass filter (d) both (a)&(b)
17. To extract deep learning features from a 3x3 image $Y=[0 \ 1 \ 1; 1 \ 0 \ 1; 1 \ 1 \ 0]$ using a convolution with a kernel of size **3x3 with all values of 1/9**, then the first output pixel is.....
- (a) 1/9 (b) 2/3 (c) 1 (d) 3
18. If a **5-fold** cross validation is applied on a data of total size of 1000 record, then the training data should be of size.....record
- (a) 800 (b) 500 (c) 200 (d) 5
19. If a **leave-one-out** cross validation is applied on a data of total size of 1000 record, then one should apply a number of experiments
- (a) 1000 (b) 999 (c) 200 (d) 1
20. To more accurately describe an accuracy to be 0.8, three experiments accuracies are....
- (a) 0.7,0.8,0.9 (b) 0.75,0.8,0.85 (c) 0.6,0.8,1 (d) any answer
21. If a blood cell is moving with a velocity u m/s towards a transducer that submits an ultrasonic wave with a frequency ν and speed C , then the reflected wave has
- (a) same speed C (b) different frequency ν' (c) different speed C' (d) both (a)&(b)
22. Continuous wave Doppler affords.....measurements of blood speed
- (a) high speed value (b) large area (c) small area (d) both (a)&(b)
23. If a blood cell is moving with a velocity 500 m/s away from a transducer that submits an ultrasonic wave with 100 MHz and 1500 m/s, then the reflected frequency is.... MHz
- (a) 75 (b) 100 (c) 150 (d) none of the given
24. If a blood cell is moving with a velocity 0 m/s and a transducer submits an ultrasonic wave with 100 MHz and 1500 m/s, then the reflected frequency is.... MHz
- (a) 75 (b) 100 (c) 150 (d) none of the given
25. Color Doppler superimpose blood velocity map on top of a(n)..... scan
- (a) A-mode (b) B-mode (c) M-mode (d) none of the given
26. If a bat hears an echo 3.4 second after it makes a chirp of speed 340 m/sec, how far away is the object?
- (a) 100m (b) 50m (c) 200m (d) 578m

MODEL BOnly answers in the electronic answer form will be graded

27. High ultrasonic imaging frequencies is suitable for imaging
- (a) small size objects (b) surface structures (c) deep structures (d) both (a)&(b)
28. In amplitude-mode ultrasonic (US) scan, the.....decreases as the amplitude increases
- (a) tissue reflectivity (b) object depth (c) transducer size (d) both (a)&(b)
29. Compound imaging (sonoCT) is an enhanced scan
- (a) A-mode (b) B-mode (c) M-mode (d) none of the given answers
30. To build a prior shape model, a registration step is done to..... training images
- (a) maximize similarities between (b) enhance visual appearance of (c) minimize similarities between (d) both (a)&(b)
31. If among 100 training images, the pixel ($x=5, y=5$) is an object for 20 images, then the **probabilistic shape value** for this pixel to **belong to the background** is
- (a) 20/100 (b) 80/100 (c) (5x5)/100 (d) none of the given
32. If the pixel ($x=5, y=5$) in the training image belong to the object edge, then the distance map **shape value** for this pixel is
- (a) 0 (b) 1 (c) -1 (d) none of the given
33. If the pixel ($x=50, y=50$) in the training image is deep inside the object, then the distance map **shape value** for this pixel is
- (a) 0 (b) positive (c) negative (d) 50x50
34. If the pixel ($x=50, y=50$) is an edge point in the training image, then the distance map **shape value** for the pixel ($x=51, y=51$) may be
- (a) 0 (b) root of 2 (c) -2 (d) none of the given
35. Forms of ionized radiation include....
- (a) MRI waves (b) ultrasonic wave (c) microwave (d) gamma-rays
36. Positron emission tomography basically used a camera to detect
- (a) positrons (b) gamma-rays (c) alpha particles (d) either (a) or (b)
37. A radioactivity is dropped by one quarter after travelling 2cm in tissue and by one more quarter after travelling a more 1cm in the tissue, then.....
- (a) half life=2cm (b) half life=3cm (c) half value layer=2cm (d) half value layer=3cm
38. Tumor therapy applications would use ...particles due its half value layer
- (a) alpha/small (b) alpha/large (c) positron/small (d) positron/large
39. A cyclotron is used onsite for PET scan as the positron-emitting radionuclide.....
- (a) has short half time (b) is synthesized (c) exists naturally (d) both (a)&(b)
40. In MRI, the RF coil.....
- (a) transmits excitation signal (b) receives body magnetic response (c) aligns protons in the filed direction (d) both (a)&(b)
41. The mass number of the following nucleus represent an MRI active nucleus:
- (a) 18 (b) 12 (c) 11 (d) either (a) or (b)